





INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁶ : G01N 33/543, C12Q 1/68	A1	(11) International Publication Number: WO 99/14596 (43) International Publication Date: 25 March 1999 (25.03.99)
(21) International Application Number: PCT/SE98/01562 (22) International Filing Date: 2 September 1998 (02.09.98) (30) Priority Data: 9703314-6 15 September 1997 (15.09.97) SE (71) Applicant (for all designated States except US): AB SANGTEC MEDICAL [SE/SE]; P.O. Box 20045, S-161 02 Bromma (SE). (72) Inventors; and (75) Inventors/Applicants (for US only): BERGGREN, Christine [SE/SE]; Iliongränden 11, S-224 72 Lund (SE). JOHANSSON, Gillis [SE/SE]; Småskolevägen 37, S-224 64 Lund (SE). (74) Agents: BERG, Sven, Anders et al.; Albihns Patentbyrå Stockholm AB, P.O. Box 3137, S-103 62 Stockholm (SE).	(81) Designated States: AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CU, CZ, DE, DK, EE, ES, FI, GB, GE, GH, GM, HR, HU, ID, IL, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, UA, UG, US, UZ, VN, YU, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG). Published With international search report.	


(54) Title: CAPACITY AFFINITY SENSOR

(57) Abstract

This invention describes a capacity affinity sensor based on self-assembled monolayers on an electrode with immobilized recognition elements available to analyte in the surrounding solution. Additional insulation is provided by auxiliary self-assembled molecules. The sensor has exceptional sensitivity and wide operating range due to these parts of the invention. It is versatile because different kinds of recognition elements can be immobilized directly to the self-assembling monolayer. The electrode then becomes selective to those molecules in the solution, the analytes, that show affinity to the recognition element on the surface. Compared to capacitive sensors described before those described here shows at least a 1000-fold better sensitivity because of the properties of the layer binding the recognition element.



(1b)



b)

Best Available Copy

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece			TR	Turkey
BG	Bulgaria	HU	Hungary	ML	Mali	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MN	Mongolia	UA	Ukraine
BR	Brazil	IL	Israel	MR	Mauritania	UG	Uganda
BY	Belarus	IS	Iceland	MW	Malawi	US	United States of America
CA	Canada	IT	Italy	MX	Mexico	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NE	Niger	VN	Viet Nam
CG	Congo	KE	Kenya	NL	Netherlands	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NO	Norway	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	NZ	New Zealand		
CM	Cameroon			PL	Poland		
CN	China	KR	Republic of Korea	PT	Portugal		
CU	Cuba	KZ	Kazakhstan	RO	Romania		
CZ	Czech Republic	LC	Saint Lucia	RU	Russian Federation		
DE	Germany	LI	Liechtenstein	SD	Sudan		
DK	Denmark	LK	Sri Lanka	SE	Sweden		
EE	Estonia	LR	Liberia	SG	Singapore		

Capacity affinity sensor

Detecting interactions between molecules forms the basis of many analytical methods. The interaction can be detected and quantified through
5 a number of schemes, e.g. precipitation, separation or through different marker molecules or reactions. Such an example is the development of immunoassays during the last three decades, which has revolutionized determination of drugs and hormones in clinical and pharmaceutical chemistry as well as contaminants in the environmental area. Almost all
10 immunomethods require labels attached either to the antibody or the antigen. Another example is the binding between a DNA-probe and its complementary DNA-strand or DNA-fraction. A number of receptors or the complementary molecule can be studied using the same approach.

15 There are a number of disadvantages associated with labels. If they are radioactive the work has to be carried out under strict safety regime and handling of waste is costly. The use of enzymes as labels requires an additional time-consuming incubation step. Common for all labels are that they require a synthetic coupling to either an antigen or an antibody or
20 generally to the recognition element or the analyte. A big label may change the affinity between the molecules which is of particular concern when an assay is performed by competition between an analyte from the sample and an added labeled molecule. Many affinity interactions cannot be studied because of this. Recognition of DNA-binding through the use of
25 electrochemical intercalators shows low sensitivity. Many attempts have therefore been made to detect the binding itself by potentiometric [Taylor, R. F.; Marenchic, I. G.; Spencer, R. H. *Anal. Chim. Acta* 1991, 249, 67-70], piezoelectric [Roederer, J. E.; Bastiaans, G. J. *Anal. Chem.* 1983, 55, 2333-

2336], or optical measurements [Löfås, S. *Pure Appl. Chem.* 1995, 67, 829-834].

Attempts have previously been made to use capacitance measurements for
5 detecting molecular interactions without the use of labels. A molecule with
affinity for the analyte should be immobilized on a conducting electrode
surface so that it can interact with the analyte in solution in such a way that
the interaction causes a change in capacitance. This principle has been used
in immunochemistry, by immobilization to oxide surfaces [Bataillard, P.;
10 Gardies, F.; Jaffrezic-Renault, N.; Martelet, C.; Colin, B.; Mandrand, B.
Anal. Chem. 1988, 60, 2374-2379] or for recognition of DNA-sequences
[Souteyrand, E.; Martin, J. R.; Cloarec, J. P.; Lawrence, M. *Eurosensors X,
The 10th European Conference on Solid-State Transducers*, 1996, Leuven,
Belgium].

15 Self-assembled monolayers of thiols, sulfides and disulfides on gold
electrodes have been widely studied and long-chain alkanethiols are known
to form insulating well-organized structures on gold substrates [Porter, M.
D.; Bright, T. B.; Allara, D. L.; Chidsey, C. E. D. *J. Am. Chem. Soc.* 1987,
20 109, 3559-3568]. The binding formed between the sulphur atom and gold is
very strong and the formed self-assembled monolayers (SAM's) are stable in
air, water and organic solvents at room temperature [Bain, C. D.;
Troughton, E. B.; Tao, Y.-T.; Evall, J.; Whitesides, G. M.; Nuzzo, R. G. *J.
Am. Chem. Soc.* 1989, 111, 321-335]. It has been suggested that
25 microcontact printing [Mrksich, M.; Whitesides, G. M. *Tibtech* 1995, 13,
228-235] and photolithography [Bhatia, S. K.; Hickman, J. J.; Ligler, F. S.
J. Am. Chem. Soc. 1992, 114, 4432-4433] can be used to pattern surfaces
with functionalized self-assembled monolayers for biosensor production

with low cost for a diversity of applications, but until now it has not been possible to produce direct affinity sensors with high sensitivity.

Terrettaz et al, *Langmuir* 1993, 9, 1361-1369, discloses a sensor, e.g. for
5 assaying cholera toxin, where the ganglioside GM1 has been bound to a SAM layer. The detection limit for capacitance measurements using the sensor is somewhere within the range from 10^{-6} to 10^{-9} M. The article states that capacitance measurements are unsuitable for assaying cholera toxin because the capacitance changes were too small, and hence, the sensitivity
10 is too low.

Self-assembled monolayers of thiols on gold, with antigenic terminating groups have been reported before, but they had coverages of only 14, 19 or 31 % for different electrodes [Taira, H.; Nakano, K.; Maeda, M.; Takagi,
15 M. *Anal. Sci.* 1993, 9, 199-206]. The lowest measured value in the article was at an antibody concentration of 10 ng/ml, which can be compared to 1 pg/ml of antigen measured with our invention (See Example 1). The higher sensitivity obtained with our electrode can be explained by that the gold surface is first covered with a self-assembled monolayer of a thiol, sulphide
20 or disulphide giving a high coverage of the surface, thereafter the recognition element is immobilized on the surface and as the last step the surface is plugged with another thiol. The saturation seems to occur at similar concentrations in the two cases if the larger bulk of the antibody compared to the antigen is taken into account. This comparison thus supports the
25 arguments given above that a dense layer is of great importance for a high sensitivity.

DNA-probes have been immobilized e. g to SiO_2 and a sensitivity of 10 ng/ml was obtained [Souteyrand, E.; Martin, J. R.; Cloarec, J. P.; Lawrence,

M. Euroensors X, The 10th European Conference on Solid-State Transducers, 1996, Leuven, Belgium].

5 A peptide bound to an alkylthiol was also immobilized as a self-assembled layer on gold, but the antibody concentration was in this case in the mg/ml range making it a less succesful sensor [Rickert, J.; Wolfgang, G.; Beck, W.; Jung, G.; Heiduschka, P. *Biosens. Bioelectron.* 1996, 11, 757-768].

One of these previous approaches are illustrated in the patent EP 244326.
10 The recognition element is bound to an insulating layer on top of a conducting substrate, the insulating layer typically being an oxide. The oxide layer has to be thick, typically 70 nm on silicon, in order to be stable and sufficiently insulating, resulting in a low sensitivity. It is difficult to obtain good surface coverage on oxides and the recognition elements are
15 not well ordered.

Rojas, M.; Königer, R.; Stoddart, F. ; Kaifer, A. ; *J. Am. Chem. Soc.* 1995, 117, 336-343 discloses an assay method for determining ferrocene in a sample using cyclodextrin. All hydroxy groups of cyclodextrin are
20 substituted by thiol groups, and the modified cyclodextrins are chemically adsorbed to a gold surface. Empty spaces on the gold surface between the adsorbed modified cyclodextrin molecules are filled with adsorbed pentanethiols. The lowest ferrocene concentration determined is 5 μ M.

25 There is always a need for improvements of analysis techniques. Especially when assaying biochemical compounds it is often necessary to be able to determine concentrations below 1 ng/ml.

Summary of the invention

It has now turned out that unexpectedly good capacity affinity sensors, suitable for determining the presence of a certain compound of interest by capacitance measurements using an electrode which can be produced by a method comprising the steps of:

- a) providing a piece, of a noble metal where said piece optionally can be a rod or, alternatively a piece of insulating material such as glass, silica or quartz, on which a noble metal is sputtered or printed;
- 10 b) providing a first SAM-forming molecule comprising a coupling group and/or an affinity group specifically binding said compound of interest;
- c) contacting the piece in step a) with the first SAM-forming molecule in step b), thereby obtaining a self-assembling monolayer on said noble metal surface;
- 15 d) in case the first SAM-forming molecule does not comprise an affinity group, contacting said self-assembling monolayer on said noble metal piece with an affinity molecule specifically binding said compound of interest, thereby coupling the affinity molecule to the self-assembling monolayer; and
- 20 e) contacting the piece obtained in step c) or d) with a second SAM-forming molecule, thereby obtaining a noble metal surface that is at least 90%, preferably at least 97 % covered with a self-assembling monolayer.

Detailed description of the invention

25

The detection limits reported in this invention are at least three orders of magnitude better than those reported previously for capacitive immunosensors and a comparison is therefore necessary in order to explain why this invention succeeds so exceptionally well. The insights behind this

invention are that the recognition layer must be thin, well-ordered and it must cover at least 90%, preferably at least 95%, more preferably at least 97%, and most preferably at least 99% of the sensor surface. In a subsequent step, any free spots between the recognition elements are

5 "plugged", i.e. covered with a second self-assembling monolayer-forming molecule, e.g. an alkanethiol comprising 3-25 carbon atoms preferably in a straight chain, after obtaining a self-assembling monolayer comprising affinity groups, thereby increasing the tightness and insulation. A capacitive biosensor is covered by an immobilized layer with the recognition element

10 toward the solution. Electrically it is equivalent to a capacitor between the conducting metal electrode and the conducting solution. Another layer forms when a molecule binds to the recognition element thereby replacing aqueous solution with a non-conducting organic molecule. This is equivalent to the formation of an additional capacitor in series with the first,

15 thereby decreasing the total capacitance.

Any part of the surface that allows the aqueous solution to penetrate below the plane where the recognition takes place will act like a short-circuiting element. The capacitance will therefore increase due to the higher dielectric

20 constant of the penetrating aqueous solution. Oxide layers are not well ordered and it is therefore impossible to form a dense recognition layer. Self-assembled monolayers are much better ordered and a more perfect coverage can therefore be expected in the immobilized layers. Furthermore the self-assembled monolayers are much thinner than the oxide layers,

25 resulting in a larger capacitance in series with the capacitance formed when molecules bind on the surface. This makes it easier to detect changes in the capacitance when an analyte binds to the surface.

This invention describes an capacity affinity sensor based on measurements of the capacitance change at conducting surfaces. The grafted recognition layer should be electrically insulating to prevent interferences from redox couples in the electrolyte solution and high Faradaic background currents.

5 On the other hand, it should be as thin as possible in order to achieve high sensitivity. The use of self-assembled binding to gold or other noble metals gives especially thin and compact layers. The invention also shows how additional insulation can be obtained by plugging with a different type of self-assembling molecule.

10

Accordingly, the present invention relates to a method for producing a capacity affinity sensor, wherein a piece of a noble metal is covered with a layer of a self-assembling monolayer-forming molecule comprising coupling groups. Affinity molecules are then coupled to these self-

15 assembling monolayer-forming molecules. Subsequently any remaining free spots on the noble metal surface is covered by a second self-assembling monolayer-forming molecule.

In another aspect, the present invention relates to a capacity affinity sensor
20 comprising a noble metal piece substantially completely covered with a self-assembling monolayer comprising first and second self-assembling monolayer-forming molecules, and where affinity molecules that specifically binds to a certain molecule of interest have been coupled to the first self-assembling monolayer-forming molecules.

25

In yet another aspect, the present invention relates to a method for qualitatively or quantitatively determining the presence of a certain compound of interest. A capacity affinity sensor, comprising a noble metal piece substantially completely covered with a self-assembling monolayer

comprising first and second self-assembling monolayer-forming molecules, and where affinity molecules that specifically binds to a certain molecule of interest have been coupled to the first self-assembling monolayer-forming molecules, is contacted with a liquid sample comprising the compound of interest and the sensor's capacitance is determined.

In a further aspect, the present invention relates to using said sensors for analysing certain compounds of interests, such as human chorionic gonadotropin hormone (HCG), interleukin-2, human serum albumin, atrazine or a certain DNA sequence.

Definitions:

As disclosed herein, the terms "self-assembled monolayer" and "SAM" are synonyms and relates to the spontaneous adsorption of film components from a solution onto a solid surface making a well-ordered monolayer. Such a layer on gold substrates have previously been described substrates [Porter, M. D.; Bright, T. B.; Allara, D. L.; Chidsey, C. E. D. *J. Am. Chem. Soc.* 1987, 109, 3559-3568].

As disclosed herein, the term "noble metal" relates to a metal chosen from the group of gold, silver, copper, platinum and palladium. Gold is preferred.

As disclosed herein, the term "affinity molecule" relates to a molecule which specifically binds to a certain molecule of interest. If the molecule to be determined is an antigen, the affinity molecule might be an antibody, preferably a monoclonal antibody, or an antibody fragment such as a F(ab')₂ fragment. If a certain nucleic acid sequence is to be identified, the affinity molecule might be a nucleic acid probe specifically hybridizing to

said nucleic acid sequence. The present invention can also be used in relation to affinity-mediating biomolecules in general, for example in situations where certain nucleic acids bind to antigens other than nucleic acids, such as proteins. The skilled person is well aware of how to choose
5 suitable affinity molecules for a certain compound to be determined.

As disclosed herein, the term SAM-forming molecule relates to a molecule having the ability of forming a self-assembling monolayer on a noble metal. A SAM-forming molecule comprises at least one thiol, sulphide or
10 disulphide group and may optionally also comprise an affinity group. Affinity molecules are coupled to small SAM-forming molecules comprising coupling groups in a separate step. Examples of such small SAM-forming molecules comprising coupling groups are thioctic acid and cysteamine. This coupling step is carried out after formation of the self-
15 assembling monolayer on the noble metal surface. The skilled person is well aware of how to choose suitable coupling reactions and coupling groups. In the following examples, a self-assembling monolayer consisting of thioctic acid is activated by 1-(3-dimethylaminopropyl)-3-ethyl-carbodiimide. Subsequently, an affinity molecule is coupled to the activated monolayer.
20 However, other similar coupling reactions are described in the literature.

As disclosed herein, the term "plugging" refers to treatment in a solution containing a thiol, sulphide or disulphide after immobilization of the affinity molecule to a self-assembling monolayer on a noble metal surface in order
25 to block any unblocked spots on said surface. As already mentioned, it is necessary that the noble metal surface is as completely covered by a SAM as possible in order to optimize the sensitivity of the sensor. Suitable examples of SAM-molecules that can be used for plugging are thiols

comprising 3 - 25 carbon atoms in a straight saturated chain. Such SAM-molecules lack coupling groups. A preferred example is 1-dodecanethiol.

As disclosed herein, SCE stands for the saturated calomel electrode;

- 5 Potentiostatic perturbation means a fast change in potential; HCG stands for human chorionic gonadotropin; IL-2 stands for interleukin 2 and HSA stands for human serum albumin.

- 10 The interactions that can be measured using this capacitance sensor includes antigen-antibody, hapten-antibody, peptide-protein, nucleic acids, lectin-hydrocarbon-containing parts, biotin-streptavidin-avidin, receptors-agonist-antagonist, ligand-cells. Complexes can be one part of the affinity pair, e. g. hapten-antibody binding to immobilized hapten as recognition element. Fragment, e. g. of antibodies can be used instead of the native antibody.
- 15 Recognition element as used in here constitutes any one of the pairs or complexes mentioned above which is immobilized on the electrode surface. Analyte is the molecule to be determined and is normally the other part than the recognition element in the pairs above.

- 20 In this invention a solution containing the molecules, complexes or cells to be determined is allowed to make contact with a surface containing the affinity group, after which the capacitance or impedance change when an interaction takes place is determined. The capacitance change takes place between the solution and a metal surface, consisting of solid metal or metal
- 25 sputtered or printed on an underlaying non-conducting surface. Faradaic reactions with the metal as well as background currents are blocked by the affinity group on the surface, eventually improved by treatment with auxiliary compounds which improve the insulation. The affinity group is bound to the metal surface, either directly through self-assembly, or by

binding it to a self-assembled compound on the electrode. It can also be bound through adsorption, polymerization or coating. Measurements are made using electrochemical perturbations followed by recording of the resulting response. The perturbations used in the examples described below
5 are potentiostatic steps or pulses which give rise to current transients from which the capacitance is evaluated. Perturbations can also be amperometric steps in which case the change in potential is used for capacitance evaluation. Perturbations with sinusoidal or other wave-forms have been reported in the literature. The sensitivity can be improved by allowing a
10 solution containing a secondary specific ligand to bind to the analyte already on the surface, thereby increasing the size of the bound aggregate and the capacitance change.

The invention will now be described in more detail with reference to the
15 enclosed drawings.

Fig. 1a shows schematically how an antibody can be immobilized to a metal surface. An alkane thiol provides additional insulation. It is also shown how the total capacitance is made up from a series connection of those of the
20 double layer, the antibody and the self-assembled layer.

Fig. 1b shows the equivalent circuit used for evaluation of the capacitance.

Fig. 2 shows the measuring flow cell, a) measuring electrode, b) auxiliary platinum foil electrode, c) platinum wire reference electrode, d) Ag/AgCl
25 reference electrode.

Fig. 3 shows the cyclic voltammetry responses in $\text{Fe}(\text{CN})_6^{3-}$ when the measuring electrode was covered with a) thiocetic acid, b) thiocetic acid and

antibody c) thioctic acid, antibody and dodecanethiol. More details are given in example 1.

Fig. 4 shows detection of human chorionic gonadotropin hormone (upper curve) and the the lower curve the absence of response to the non-specific thyrotropic hormone (lower curve) as specified in example 1.

Fig. 5 shows that $F(ab')_2$ fragments can be used as recognition elements for the human chorionic gonadotropin hormone as described in example 2.

10

Fig. 6 shows that reduced $F(ab')_2$ fragments can be used as recognition elements for the human chorionic gonadotropin hormone as described in example 3.

15 Fig. 7 shows detection of the cytokine Interleukin-2 as mentioned in example 4.

Fig. 8 shows detection of human serum albumin in a flow cell with different flow rates as discussed in example 5.

20

Fig. 9 shows the structure of the modified atrazine discussed in example 6.

Fig. 10 shows binding of antibodies to atrazines with different side arms, as discussed in example 6.

25

Fig. 11 shows the binding of a cytomegalo virus single stranded 179 base DNA-fragment to an 8 bases long recognition element on the measuring electrode (upper curve) and the non-specific control with a single-stranded

207 base DNA fragment from tyrosinase (lower curve). See example 7 for details.

Fig. 12 shows the binding of a cytomegalo virus single stranded 179 base
5 DNA-fragment to a 25 bases long recognition element on the measuring
electrode. See example 8 for details.

Fig. 13 shows the binding of a cytomegalo virus single stranded 179 base
DNA-fragment to a 25 bases long recognition element immobilized so it
10 will lie flat on the surface of the measuring electrode (upper curve) and the
control without recognition element (lower curve).

If a solid measuring metal electrode is used, a gold rod typically 3 mm in
diameter, is polished, cleaned and coated through self-assembly with a
15 recognition element or with a compound which can be coupled with a
recognition element. A great number of coupling methods are known and
may be used as alternatives to those described in the examples. It is also
possible to use metal sputtered or printed on glass, quartz, silicon or another
insulating materials as disposable electrodes. After cleaning the electrodes
20 are coated in batch and inserted in a quick-connect measuring cell. A
number of different recognition elements can be put on the same sputtered
electrode if they are separated by insulating parts and connected to the
potentiostat with switches which can be controlled by a microprocessor.

25 The importance of making the recognition layer thin and with a large
capacitance is illustrated by Fig. 1 with a coupling chemistry as in Example
1. The inverse total capacitance is the sum of the inverse capacitances of
each layer in series, i. e. the thioctic acid layer, the antibody layer and the
capacitance between the antibody and solution. If one of these is small

compared to the others, it will dominate the total capacitance. Specially if self-assembled parts give rise to a small capacitance, it will dominate over the capacitances in the recognition layer. Changes in the recognition layer will thus have little effect on the total capacitance resulting in a low overall sensitivity of the sensor.

The electrode is inserted into a cell which may be either of the cuvette type or a flow cell as shown in Fig. 2. The cell must contain an auxiliary electrode, typically a platinum foil which should be placed symmetrically and opposite to the measuring electrode. A reference electrode, typically SCE, is placed in the cell so that the voltage drop between the reference and measuring electrodes due to capacitive or Faradaic currents becomes very small. In some cases the performance may be improved if a very small additional reference electrode is used, see Fig. 1c and the SCE reference is moved away, Fig. 1d. A flow cell gives more precise control over the mass transfer to the measuring electrode and injection of sample and cleaning up is more easily automated. Flow cells with volumes of 2 ml and 10 μ l were found to have about the same sensitivity. A flow cell with disposable electrodes made by sputtering gold on silicon also had similar properties.

20

The electrodes are connected to a fast potentiostat which in turn is controlled from a microprocessor. The potentiostat will keep the measuring electrode at a pre-set value *versus* the reference. A potentiostatic perturbation is imposed on the measuring electrode. The currents caused by the perturbation voltage are used for evaluation of the capacitance of the measuring electrode.

25

A known volume of sample is normally mixed with a known volume of a conducting liquid in a cuvette in a batch cell. In the case of a flow cell a

known volume is injected into a conducting carrier flow pumped with a known flow rate. The conducting liquids are normally buffers with ionic strengths from a few millimolar and up. The sample can be in a non-conducting medium but a conducting solution must fill the cell when
5 measurements are made.

The invention will now be further described in the following examples. These examples are given for the purpose of illustration and are not intended to limit the scope of the invention.

10

Example 1.

The antibody-covered electrode is schematically shown in Figure 1 a. The electrode was a gold rod (99.99% Aldrich, 3 mm in diameter) cut up into
15 thin sections threaded to stainless steel holders. Prior to immobilization the gold rod was polished with alumina slurries down to 0.04-0.05 μm . After mounting into the Teflon holder the electrode was plasma cleaned for 15 min and immediately placed in a solution of 2 % (w/w) D/L-thioctic acid in absolute ethanol. The electrode was taken from the solution after 24 hours,
20 thoroughly rinsed in absolute ethanol and allowed to dry. Thereafter the electrode was put into a solution of 1 % (w/w) 1-(3-dimethylaminopropyl)-3-ethyl-carbodiimide hydrochloride in dried acetonitrile for 5 hours. 5 μl (approximately 1 mg/ml) antibody solution was placed on the electrode surface and the coupling procedure was performed at 4 $^{\circ}\text{C}$ for 24 hours. The
25 coupling procedure followed essentially that described by Duan *et al* [Duan, C.; Meyerhoff, M. E. *Anal. Chem.* 1994, 66, 1369-1377]. A long thiol, 1-dodecanethiol was used to "plug", i.e. block any unblocked spots on the electrode surface.

Capacitance measurements.

The capacitance changes were evaluated from the transient current response obtained when a potentiostatic step was applied to the electrode. An
5 alternative measurement principle relies on the evaluation of the currents at a number of sinusoidal wave frequencies, usually called impedance spectroscopy. The two methods have been compared using the same potentiostat and electrode and found to give almost the same results in terms of equivalent capacitances and resistances. The potentiostatic step method is
10 faster and more convenient and is therefore used here.

The measuring set-up consisted of a three-electrode system, with an extra reference electrode, connected to a fast potentiostat. The potentiostat was connected to a computer (486, 33 MHz) via a Keithley 575 measurement
15 and control system, containing 16-bit A/D and D/A converters. The Keithley system was powered from the computer through a galvanically isolated power line in the box. The potentiostat was powered from the Keithley in order to isolate the analog parts from the noisy digital circuits. The sampling frequency of 50 kHz was determined by an internal clock in
20 the Keithley box. The current values were taken as the mean of ten repeated steps. The rest potential was 0 mV vs. an Ag/AgCl reference electrode. A potential step of 50 mV was applied and the current transient that followed was sampled. An identical current transient but of opposite direction was obtained when the potential was stepped back to the rest value.

25

Taking the logarithm of the current gives an almost linear curve from which R_s and C_1 can be calculated (see Figure 1 b) using the equation:

$$i(t) = u/R_s \exp(-t/R_s * C_1)$$

where $i(t)$ is the current in the circuit as a function of time; u is the applied pulse potential; R_s is the resistance between the gold surface and the reference electrode; t is the time elapsed after the potentiostatic pulse was applied and C_1 is the capacitance measured between the gold electrode and the solution. The first ten current values were used for the calculation and a correlation coefficient of better than 0.99 was obtained.

A platinum wire was used as a reference electrode because it can be placed closer to the working electrode than a Luggin capillary of glass without causing any shielding. This will sharpen the current transient and improve the accuracy of the measurements. The platinum reference electrode, though, has no defined potential so its potential was compared to a commercial Ag/AgCl reference electrode, Fig. 1d, just before the potentiostatic pulse was applied.

The carrier solution, 10 mM citrate buffer, pH 7.4 was pumped with a flow rate of approximately 0.5 ml/min through the flow cell. An injector with a loop of 250 μ l was connected to the flow system.

Cyclic voltammetry.

Cyclic voltammograms were recorded in a three-electrode system in a batch cell. The working electrode was the unmodified or modified gold rod (3 mm in diameter) in a Teflon holder, the auxiliary electrode was a platinum foil and the reference electrode was a saturated calomel electrode (SCE). 5 mM of a $K_3(Fe(CN)_6)$ solution was used for the measurements. The instrumentation used for cyclic voltammetry was a Princeton Applied 273 A potentiostat controlled by a computer.

A gold surface covered with a long chain alkanethiol layer blocks almost all faradaic currents and is highly insulating with an equivalent transfer and dynamic resistance of about 2 000 and $69 \Omega\text{cm}^2$, respectively for a surface covered with butanethiol [Swietlow, A.; Skoog, M.; Johansson, G. *Electroanal.* 1992, 4, 921-928]. A layer of thioctic acid was much less insulating with an equivalent transfer and dynamic resistance of 470 and $40 \Omega\text{cm}^2$, respectively. The permeability of ions through the layer is so high that a redox couple can penetrate it, giving almost the same currents in a cyclic voltammogram as on a bare gold electrode, see Fig. 3, curve a. Immobilization of a monoclonal antibody towards human chorionic hormone (HCG) reduces the penetration of the redox couple, Fig. 3, curve b. Insulation is further improved when the electrode is treated with 1-dodecanethiol as can be seen from the absence of redox peaks for such an electrode, Fig. 3, curve c.

Antigen detection.

When an antigen binds to the antibody immobilized on the electrode, there will be an additional layer decreasing the total C_1 further. The binding between the antigen and antibody is therefore detected directly. No label is necessary for the antigen. The physical basis for the response is thought to arise from displacement of the polar water further out from the electrode surface replacing it with a much less polar molecule.

25

The human chorionic gonadotropin hormone, HCG, was used as model substance. HCG is a glycoprotein with a molecular weight of 30 000 D. The hormone consists of an alpha and a beta chain. The alpha chain is the same as in the thyrotropic hormone, but the beta chains differ in the two

hormones. The monoclonal antibody immobilized on the electrode was directed towards the beta chain specific for HCG. Thyrotropic hormone and HCG are known to have a cross-reactivity of less than 0.05 % [Sigma Chemical Co., Product specification, C-7659]. The thyrotropic hormone
5 was used as a control for testing the selectivity of the immunosensor.

Samples with HCG-concentrations as low as $30 \cdot 10^{-15}$ M (1 pg/ml) were injected into the flow system. The capacitance was continuously measured and found to decrease after an injection until it reached a stable value,
10 which took approximately 15 minutes in the 2 ml cell with a flow rate of 0.5 ml/min. The change in capacitance vs. the logarithm of the concentration, was found to give a linear relationship up to a concentration of approximately 10^{-11} M (0.3 ng/ml) and to reach a saturating value at 10^{-10} M, see Fig. 4. The detection limit was around $15 \cdot 10^{-15}$ M (0.5 pg/ml)
15 hormone. It was calculated from a comparison between the signal and the irreproducibility of measurements on the antibody surface alone. The irreproducibility corresponds to 15 nFcm^{-2} .

As usual in flow injection analysis, the sensitivity and detection limit can be
20 changed by changing the injection volume. A larger sample size will thus decrease the detection limit in proportion.

No cross-reactivity whatsoever was observed on the capacity affinity sensor, when the control antigen, thyrotropic hormone, was injected into the
25 flow system. This suggests that the observed capacitance change is specific and not caused by an unspecific adsorption of protein to the sensor surface. Injection of a serum sample without added HCG produced a 13% increase in the capacitance when the sample entered the cell. The signal returned to the previous value when buffer filled the cell again. The increase in

capacity is due to the increased ionic strength of the solution. The experiment thus shows that serum as such does not give rise to any permanent change.

5 Example 2.

Capacitance changes for antibody fragments.

To increase the sensitivity of the signal, antibodies against HCG were
10 digested with Ficin to $F(ab')_2$ fragments. The idea is to remove an inactive part of the antibody and to move the binding sites closer to the electrode surface. The fragments were immobilized to the electrode surface in the same way as described above. The analytical properties were similar to those obtained with electrodes covered with the native antibody, as shown
15 in Fig. 5. The capacitance, C_1 , of the $F(ab')_2$ electrode was 4500 nFcm^{-2} compared to 1400 nFcm^{-2} for an electrode with a native antibody. The resistivities were about $63 \Omega\text{cm}^2$ in both cases. The slopes of the calibration curves were about the same in both cases and an increased sensitivity was not obtained. The increased capacitance will improve the signal-to-noise
20 ratio somewhat.

Example 3.

The $F(ab')_2$ - fragment of HCG was reduced in 0.1 M phosphate buffer, pH
25 6, containing 0.15 M NaCl, 5 mM EDTA, 4.2 mg/ml 2-mercaptoethylamine during 1.5 h at 37 °C. The solution was ultrafiltered on an Amicon dialysis filter, cut-off 10 000 D. The plasma-cleaned gold electrode was dipped into the filtrate at room temperature over night. The electrode was later treated with 1-dodecanethiol.

The procedure illustrates a direct binding between the sulfur atom of a univalent antibody fragment and the metal. The surface will be even more homogeneous with this procedure and the antibodies' binding part is directed out into solution. The capacitance will be even higher with this treatment and the sensitivity will be higher as shown in Fig. 6.

Example 4.

A monoclonal antibody towards Interleukin-2, IL-2 (M_w 15 700 D) was immobilized as described above. The results, see Fig. 7, indicate that the capacitance change was about half as large for IL-2 as for HCG. This can be explained by the larger molecular weight of HCG.

The IL-2 antibody was taken from a commercial sandwich ELISA kit for determination of IL-2 after incubation in micro titer plates with a stated detection limit of 6 pg/ml in medium and 10 pg/ml in serum. The detection limit for the immunosensor is better than 1 pg/ml. Serum samples from apparently healthy donors were all below 31 pg/ml [R & D Systems, Inc., Quantikine, IL-2 manual], i.e. the commercial kit could not reliably measure IL-2-levels in healthy individuals.

Example 5.

A monoclonal antibody towards human serum albumin, HSA, (M_w 69 000 D) was immobilized on the electrode as described above. The response for HSA was lower than for IL-2, which suggests that more factors than molecular size has to be taken into account. Such factors can be the structure of the antigen, that is if it has a compact or a loose configuration, charges of the antigen and the affinity constant for the antibody-antigen

complex. One possibility is that albumin is penetrated by the aqueous phase resulting in an increased polarity of the antigen layer. Another possibility is that the antigen binds in such a way that aqueous solution can penetrate between molecules to some extent. There might be sterical hindrance for
5 two large HSA molecules to bind to an antibody with about the same molecular weight as the two.

The capacitance changes obtained for different flow rates were studied for the HSA system and the results are shown in Fig. 8. The capacitance change
10 was found to increase from a flow rate of 0.6 ml/min down to 0.15 ml/min. A longer residence time in the cell will allow more HSA molecules to be transported up to the sensor surface by diffusion and hydrodynamic movements in the solution. An increased sensitivity with decreasing flow rate is therefore generally expected.

15 A closer look at the curve shapes for HSA at 0.3 and 0.6 ml/min show that they differ from those of the other antigens and from that of HSA at 0.15 ml/min. The lower flow rate gives HSA more time to interact with the antibody and to rearrange itself on the sensor surface. The sensitivity per
20 molecule seems also to increase when the concentration decreases.

Example 6

The herbicide atrazine is a small molecule and the capacitance changes will
25 be small if it binds to an antibody on the measuring electrode. A competitive assay can be made by binding a bulky molecule to the herbicide and to allow this labeled antigen to compete with analyte antigens. A displacement assay can also be performed thus dispensing with the need to use labels. In this assay the antigen was bound to cysteamine self-assembled

on gold by coupling to a carboxylic group in the modified antigen with 1-(3-dimethylaminopropyl)-3-ethyl-carbodiimide hydrochloride in dried acetonitrile for 5 hours. Three different side-chains in the antigen were tested, see Fig. 9. The different modified antigens were tested by injecting
5 antibodies, see Fig. 10. It can be seen that the t-butyl derivative binds more efficiently to the antibody than the others. It saturates and reaches a constant level at low antibody concentrations. With an antibody saturated surface, addition of analyte antigen will cause the antibody to be displaced to some degree, proportional to the concentration, to form a soluble
10 complex. The capacitance will increase when the amount of antibody on the surface diminishes. There should be room for the hypervariable region of the antibody to interact with the bound antigen. If the antigens are packed too densely they may be interspaced with some inactive compounds.

15 Example 7

DNA can be detected by binding a single-stranded DNA-probe to the measuring electrode. The gold surface was treated prior to immobilization as described above. Thereafter it was placed in a thiol solution of 2 %
20 (w/w) of cysteamine in ethanol for 16 hours. After reaction the electrode was thoroughly rinsed in ethanol and dried. The coupling of the oligonucleotide to the phosphorylated 5' end was performed in an 0.1 M imidazole buffer, pH 6-7, containing 0.15 M 1-(3-dimethylaminopropyl)-3-ethyl-carbodiimide hydrochloride, at room temperature for 16 hours. After
25 reaction the electrode was rinsed in buffer and placed in the flow-cell.

An oligonucleotide consisting of 8 bases (SEQ.ID.NO.1) displaying the base sequence of the cytomegalo virus showed capacitance changes when an 179 long single-stranded DNA-fragment was injected and hybridized on

the surface, see Fig. 11. The figure also shows the result when a control consisting of a 207-base pair long single-stranded fragment from tyrosinase mRNA was used as sample. The selectivity is indeed very good.

5 Example 8

An oligonucleotide probe comprising 8 nucleotides might bind, at least with some of the bases, to sequences which occur randomly in a mixed biological sample. Another probe consisting of 20 base-pairs
10 (SEQ.ID.NO.2) was therefore immobilized on the measuring electrode in the same way as described above. The probe was towards the end of the cytomegalo virus fragment. Fig. 12 shows that a response indeed is obtained.

15 Example 9

There might be some disadvantages with probes directed towards the end of a DNA fragment. It was found, however, that with a probe directed towards a middle section the capacitance change did indeed occur at first but the
20 capacitance returned to the original value after some time in the flow. The probe was therefore immobilized so that it would lie flat on the measuring electrode surface.

34 μ l of the oligonucleotide 25-mer (SEQ.ID.NO.3) was incubated on ice
25 for 10 min. with 20 μ l 1 M NaHCO_3 , pH 9.6, 2 μ l 8 mM N-bromosuccinimide in water, and water to a final volume of 200 μ l.

Thereafter an electrode pretreated with cysteamine, as described above, was dipped in the solution and the reaction took place at 50°C during 1 hour.

Fig 13 shows the response obtained with this electrode. An electrode covered with cystamine only served as a control. No response whatsoever was obtained for the control when the 179 base pair single-stranded DNA fragment was injected.

SEQUENCE LISTING

(1) GENERAL INFORMATION:

(i) APPLICANT:

- (A) NAME: AB SANGTEC MEDICAL
- (B) STREET: P.O. BOX 20045
- (C) CITY: BROMMA
- (E) COUNTRY: SWEDEN
- (F) POSTAL CODE (ZIP): 161 02
- (G) TELEPHONE: +46-8-635 12 00

(ii) TITLE OF INVENTION: CAPACITY AFFINITY SENSOR

(iii) NUMBER OF SEQUENCES: 3

(iv) COMPUTER READABLE FORM:

- (A) MEDIUM TYPE: Floppy disk
- (B) COMPUTER: IBM PC compatible
- (C) OPERATING SYSTEM: PC-DOS/MS-DOS
- (D) SOFTWARE: PatentIn Release #1.0, Version #1.30 (EPO)

(2) INFORMATION FOR SEQ ID NO: 1:

(i) SEQUENCE CHARACTERISTICS:

- (A) LENGTH: 8 base pairs
- (B) TYPE: nucleic acid
- (C) STRANDEDNESS: single
- (D) TOPOLOGY: linear

(ii) MOLECULE TYPE: DNA (genomic)

(iii) HYPOTHETICAL: NO

(iv) ANTI-SENSE: NO

(xi) SEQUENCE DESCRIPTION: SEQ ID NO: 1:

TTAGGAGA

8

(2) INFORMATION FOR SEQ ID NO: 2:

(i) SEQUENCE CHARACTERISTICS:

- (A) LENGTH: 20 base pairs
- (B) TYPE: nucleic acid
- (C) STRANDEDNESS: single
- (D) TOPOLOGY: linear

27

(ii) MOLECULE TYPE: DNA (genomic)

(iii) HYPOTHETICAL: NO

(iv) ANTI-SENSE: NO

(xi) SEQUENCE DESCRIPTION: SEQ ID NO: 2:

TAGGGAAGGC TGAGTTCTTG

20

(2) INFORMATION FOR SEQ ID NO: 3:

(i) SEQUENCE CHARACTERISTICS:

(A) LENGTH: 25 base pairs

(B) TYPE: nucleic acid

(C) STRANDEDNESS: single

(D) TOPOLOGY: linear

(ii) MOLECULE TYPE: DNA (genomic)

(iii) HYPOTHETICAL: NO

(iv) ANTI-SENSE: NO

(xi) SEQUENCE DESCRIPTION: SEQ ID NO: 3:

TAGGGAAGGC TGAGTTCTTG GTAAA

25

Claims

- 5 1. A method for producing a capacity affinity sensor, suitable for determining the presence of a certain compound of interest by capacitance measurement, comprising the steps of:
- a) providing a piece of a noble metal where said piece optionally can be a rod, or alternatively a piece of insulating material such as glass, silicon,
10 or quartz, on which a noble metal is sputtered or printed;
- b) providing a first SAM-forming molecule comprising a coupling group;
- c) contacting the piece in step a) with the first SAM-forming molecule in step b), thereby obtaining a self-assembling monolayer on said noble metal surface;
- 15 d) contacting said self-assembling monolayer on said noble metal piece with an affinity molecule specifically binding said compound of interest, thereby coupling the affinity molecule to the self-assembling monolayer;
- e) contacting the piece obtained in step c) or d) with a second SAM-forming molecule, thereby obtaining a noble metal surface that is at least 90%,
20 preferably at least 95 %, morepreferably at least 97 %, and most preferably at least 99 % covered with a self-assembling monolayer.
2. A method according to claim 1, characterized in that the coupling group of the first SAM-forming molecule is activated before step d) is carried out.
- 25 3. A method according to claim 2, characterized in that the first SAM-forming molecule is D/L-thioctic acid and in that this molecule is activated with 1-(3-dimethylaminopropyl)-3-ethyl-carbodiimide.

4. A method according to claim 1, characterized in that the second SAM-forming molecule is a thiol comprising 3 - 25 carbon atoms in a straight saturated chain, and preferably is 1-dodecanethiol.
- 5 5. A capacity affinity sensor comprising a piece of a noble metal, where said piece optionally can be a rod, or alternatively a piece of insulating material such as glass, silicon or quartz on which a noble metal is sputtered or printed, to which noble metal surface affinity groups have been bound, characterized in that it has been produced by a method according to
10 anyone of claims 1-4, and where said affinity groups are comprised in a self-assembling monolayer covering at least 90%, preferably at least 95%, more preferably at least 97%, and most preferably at least 99% of the surface of the noble metal piece.
- 15 6. A sensor according to claim 5, characterized in that said affinity groups are antibodies, preferably monoclonal antibodies, or antibody fragments, preferably $F(ab')_2$ fragments.
- 20 7. A sensor according to claim 5, characterized in that said affinity groups are nucleic acid molecules, preferably single-stranded DNA molecules.
8. A method for qualitatively or quantitatively determining the presence of a certain compound of interest in a liquid sample, comprising the steps of:
a) providing a sensor according to claim 5, wherein said affinity groups
25 specifically binds to said compound of interest;
b) contacting said sensor with a reference liquid not containing said compound and determining the capacitance according to per se known methods;

- c) contacting said sensor with a sample suspected of containing said compound and determining the capacitance according to per se known methods; and
- d) calculating the difference between the capacitance of the sample and the capacitance of the reference, and optionally calculating the amount of said compound by using prerecorded calibration data.

9. A method according to claim 8 for determining the presence of human chorionic gonadotropin hormone (HCG), interleukin-2, human serum albumin, atrazine, or a certain DNA sequence.

10. Use of a sensor according to claim 5 for determining the presence of human chorionic gonadotropin hormone (HCG), interleukin-2, human serum albumin, atrazine, or a certain DNA sequence.

1
INTERNATIONAL SEARCH REPORT

International application No.

PCT/SE 98/01562

A. CLASSIFICATION OF SUBJECT MATTER

IPC6: G01N 33/543, C12Q 1/68

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC6: G01N, C12Q

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

SE,DK,FI,NO classes as above

Electronic data base consulted during the international search (name of data base and, where practicable, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	J. Am. Chem. Soc., Volume 117, 1995, María T. Rojas et al, "Supported Monolayers Containing Preformed Binding Sites. Synthesis and Interfacial Binding Properties of a Thiolated Beta-Cyclodextrin Derivative" page 336 - page 343	1,2,4,5,8
Y	--	1-10
Y	Langmuir, Volume 9, 1993, Samuel Terrettaz et al, "Protein Binding to Supported Lipid Membranes: Investigation of the Cholera Toxin-Ganglioside Interaction by Simultaneous Impedance Spectroscopy and Surface Plasmon Resonance" page 1361 - page 1369	1-10
	--	

☒ Further documents are listed in the continuation of Box C. ☒ See patent family annex.

- * Special categories of cited documents:
- "A" document defining the general state of the art which is not considered to be of particular relevance
- "E" earlier document but published on or after the international filing date
- "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- "O" document referring to an oral disclosure, use, exhibition or other means
- "P" document published prior to the international filing date but later than the priority date claimed
- "T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
- "X" document of particular relevance: the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
- "Y" document of particular relevance: the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art
- "&" document member of the same patent family

Date of the actual completion of the international search

Date of mailing of the international search report

16 December 1998

29 -12- 1998

Name and mailing address of the ISA/
Swedish Patent Office
Box 5055, S-102 42 STOCKHOLM
Facsimile No. +46 8 666 02 86

Authorized officer

Carl-Olof Gustafsson
Telephone No. +46 8 782 25 00

INTERNATIONAL SEARCH REPORT

International application No.

PCT/SE 98/01562

C (Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y	J. Am. Chem. Soc., Volume 113, 1991, Suzi Steinberg et al, "Ion-Selective Monolayer Membranes Based upon Self-Assembling Tetradentate Ligand Monolayers on Gold Electrodes. 2. Effect of Applied Potential on Ion Binding" page 5176 - page 5182 --	1-10
Y	Analytical Sciences, Volume 9, April 1993, Hiroaki Taira et al, "Electrode Modification by Long-Chain, Dialkyl Disulfide Reagent Having Terminal Dinitrophenyl Group and Its Application to Impedimetric Immunosensors" page 199 - page 206 --	1-10
Y	US 5466348 A (JAMES W. HOLM-KENNEDY), 14 November 1995 (14.11.95), column 23 - column 24 --	1-10
Y	WO 9610178 A1 (PHARMACIA BIOSENSOR AB), 4 April 1996 (04.04.96) --	1-10
Y	Analytical Chemistry, Volume 66, No 9, May 1994, Chuanming Duan et al, "Separation-Free Sandwich Enzyme Immunoassays Using Microporous Gold Electrodes and Self-Assembled Monolayer/Immobilized Capture Antibodies", page 1369 - page 1377, see page 1371-1372 --	1-10
A	US 5436170 A (BRUCE A. CORNELL ET AL), 25 July 1995 (25.07.95), see example 7 --	1
A	J. Am. Chem. Soc., Volume 117, 1995, Joseph B. Schlenoff et al, "Stability and Self-Exchange in Alkanethiol Monolayers", page 12528 - page 12536, see page 12536, left column --	1-10

INTERNATIONAL SEARCH REPORT

International application No.

PCT/SE 98/01562

C (Continuation). DOCUMENTS CONSIDERED TO BE RELEVANT

Category*	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	US 4822566 A (ARNOLD L. NEWMAN), 18 April 1989 (18.04.89), see column 9 -- -----	1-10

Form PCT/ISA/210 (continuation of second sheet) (July 1992)

INTERNATIONAL SEARCH REPORT
Information on patent family members

01/12/98

International application No.
PCT/SE 98/01562

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 5466348 A	14/11/95	AU 2907092 A CA 2121797 A GB 2278235 A,B GB 2294808 A,B GB 9410175 D GB 9600475 D WO 9308464 A	21/05/93 29/04/93 23/11/94 08/05/96 00/00/00 00/00/00 29/04/93
WO 9610178 A1	04/04/96	EP 0784793 A JP 10507126 T SE 9403245 D	23/07/97 14/07/98 00/00/00
US 5436170 A	25/07/95	AT 113724 T AU 2127988 A CA 1335879 A DE 3852036 D,T EP 0382736 A,B SE 0382736 T3 JP 2682859 B JP 3503209 T US 5693477 A US 5741712 A US 5766960 A WO 8901159 A	15/11/94 01/03/89 13/06/95 09/03/95 22/08/90 26/11/97 18/07/91 02/12/97 21/04/98 16/06/98 09/02/89
US 4822566 A	18/04/89	EP 0363401 A JP 2504069 T WO 8809499 A CA 1259374 A EP 0245477 A JP 63501446 T WO 8703095 A	18/04/90 22/11/90 01/12/88 12/09/89 19/11/87 02/06/88 21/05/87

1/12

Fig. 1a






Fig. 1b



b)


2/12

Fig.2



3/12


Fig. 3



SUBSTITUTE SHEET (RULE 26)


4/12

Fig. 4



5/12


Fig. 5



SUBSTITUTE SHEET (RULE 26)

6/12


Fig. 6



SUBSTITUTE SHEET (RULE 26)

7/12


Fig. 7



SUBSTITUTE SHEET (RULE 26)


8/12

Fig. 8



9/12


Fig. 9



R''= Et, iPr, tBu (Tracer)


10/12

Fig. 10



11/12


Fig. 11



SUBSTITUTE SHEET (RULE 26)

12/12

Fig. 12



**This Page is Inserted by IFW Indexing and Scanning
Operations and is not part of the Official Record**

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☒ **BLACK BORDERS**
- ☐ **IMAGE CUT OFF AT TOP, BOTTOM OR SIDES**
- ☐ **FADED TEXT OR DRAWING**
- ☐ **BLURRED OR ILLEGIBLE TEXT OR DRAWING**
- ☐ **SKEWED/SLANTED IMAGES**
- ☐ **COLOR OR BLACK AND WHITE PHOTOGRAPHS**
- ☐ **GRAY SCALE DOCUMENTS**
- ☐ **LINES OR MARKS ON ORIGINAL DOCUMENT**
- ☐ **REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY**
- ☐ **OTHER: _____**

IMAGES ARE BEST AVAILABLE COPY.

As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.

BEST AVAILABLE COPY

BEST AVAILABLE COPY

THIS PAGE BLANK (USPTO)